



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/646,486	08/22/2003	Eric Ryba	11350.23	9316
23862	7590	11/17/2004	EXAMINER	
NYDEGGER & ASSOCIATES 348 OLIVE STREET SAN DIEGO, CA 92103			ROLLINS, ROSILAND STACIE	
			ART UNIT	PAPER NUMBER
			3739	
DATE MAILED: 11/17/2004				

Please find below and/or attached an Office communication concerning this application or proceeding.

# Office Action Summary

Application No.

10/646,486

Applicant(s)

RYBA, ERIC

Examiner

Rosiland S Rollins

Art Unit

3739

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

## Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

## Status

- 1) ☒ Responsive to communication(s) filed on 23 August 2004.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

## Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-20 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

## Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

## Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
- ☐ Certified copies of the priority documents have been received.
  - ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  - ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

## Attachment(s)

- ☐ Notice of References Cited (PTO-892)
- ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_
- ☐ Notice of Informal Patent Application (PTO-152)
- ☐ Other: \_\_\_\_\_

## DETAILED ACTION

### *Claim Rejections - 35 USC § 103*

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

Claims 1-20 are rejected under 35 U.S.C. 103(a) as being unpatentable over Lalonde et al. (US 6283959) further in view of Benson (US 4082096) and Dobak, III (US 6482226). Lalonde et al. disclose a device for cryoablating exposed tissue comprising a tube-shaped shaft (10), a flexible enclosure (203) and a means for cooling (201) the enclosure to cryoablate tissue. Lalonde et al. '959 teach all of the limitations of the claims except a shapeable element attached to the distal end of the shaft and extending therefrom into the cryochamber and a high pressure tube with a capillary tube connected in fluid communication with the distal end of the high pressure tube.

Benson discloses a cryoablation device that includes a shapeable element that helps to maintain the desired configuration of the tip and provide a means of pressing the probe tip into engagement with tissues. Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to provide a shapeable element on the Lalonde et al. device as taught by Benson to help maintain the desired configuration of the tip.

Dobak, III disclose a similar apparatus and teach that it is old and well known to include a high pressure tube with a capillary tube connected in fluid communication with

Art Unit: 3739

the distal end of the high pressure tube since it allows cooling to be focused at the flexible enclosure and eliminates cooling along the catheter body (col. 4 lines 27-39; col. 6 lines 9+). Therefore, it would have been obvious to one having ordinary skill in the art at the time the invention was made to provide a supply tube with a coaxial capillary tube extending distally therefrom on the Lalonde et al. device, particularly in view of the teaching of Dobak, III.

### ***Response to Arguments***

Applicant's arguments filed 8/23/04 have been fully considered but they are not persuasive. In response to applicant's argument that there is no suggestion to combine the references, the examiner recognizes that obviousness can only be established by combining or modifying the teachings of the prior art to produce the claimed invention where there is some teaching, suggestion, or motivation to do so found either in the references themselves or in the knowledge generally available to one of ordinary skill in the art. See *In re Fine*, 837 F.2d 1071, 5 USPQ2d 1596 (Fed. Cir. 1988) and *In re Jones*, 958 F.2d 347, 21 USPQ2d 1941 (Fed. Cir. 1992). In this case, the motivation was found in the Benson reference as explained in the rejection above. Moreover, the argument that the combination would make certain elements of Lalonde et al. functionally redundant does not preclude the fact that the references could be combined to enhance the function in which Applicant believes is being duplicated.

### ***Conclusion***

Art Unit: 3739

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than **SIX MONTHS** from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Rosiland S Rollins whose telephone number is 703/3082711. The examiner can normally be reached on Mon.-Fri. 9:00 AM - 6:00 PM.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Linda C. Dvorak can be reached on 703/3080994. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Art Unit: 3739

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

  
Rosiland S Rollins  
Primary Examiner  
Art Unit 3739

RR